## **Certificate of Analysis**



Product Name : Water LC-MS Date Of Manufacturing : JANUARY 2024

CAS Number : 7732-18-5 Batch No. :080124

Molecular Formula: H2O Ex. Date: DECEMBER 2024

Molecular Weight : 18.01 g/mol

SR NO.	TEST	RESULT	SPECIFICATION		
	Chemical Residues in Water				
1	CAluminium (as Al), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
2	Berium (as Ba), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
3	Chromium (as Cr), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
4	Cadmium (as Cd), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
5	Cobalt (as Cu), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
6	Copper (as Co), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
7	Silver (as Ag), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
8	Zinc (as Zn), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
9	Lead (as Pb), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
10	Manganese (as Mn), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
11	Nickle (as Ni), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
12	Sodium (as Na), ppb	B.L.Q. (Q.L=1)	≤ 20 ppb		
13	Calcium (as ca), ppb	B.L.Q. (Q.L=1)	≤ 20 ppb		
14	Potassium (as Ka), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
15	Tin (as Sn), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		
16	Magnesium (as Mg), ppb	B.L.Q. (Q.L=1)	≤ 10 ppb		

Continue.....



SR NO.	TEST	RESULT	SPECIFICATION	
Water				
17	Protease	Not Detected	-	
18	Fluorescence Trace Impurities In ppb	< 0.1 ppb	< 0.1 ppb	
	At 450 nm Emission			
	At Emission Max Of Impurities	< 0.2 ppb	< 0.2 ppb	
19	Resideu after Evaporation on	< 1 ppm	≤ 1 ppm	
	heating at 110 Deg C mg/l			
20	LC-MS Gradient Suitability	Passes The Test	BY LCMSMS	
21	LC-MS at Positive Mode as	< 20 ppb	< 50 ppb	
	Reserpine			
22	LC-MS at Negative Mode as	< 20 ppb	< 50 ppb	
	Aldicarb			
23	LC Graident Suitability at 254	Passes The Test	BY LCMSMS	
	and 205 nm			
24	UV Absorption at 280nm	< 0.002	≤ 0.005	
25	UV Absorption at 260nm	< 0.002	≤ 0.005	
26	UV Absorption at 254nm	< 0.002	≤ 0.005	
27	UV Absorption at 240nm	< 0.002	< 0.01	
28	UV Absorption at 230nm	0.003	< 0.01	
29	UV Absorption at 220nm	0.004	< 0.01	
30	UV Absorption at 210nm	0.004	< 0.01	

- For laboratory, reserch or manufacturing use
- May change over time due to extraction from container
- Filtered through 0.1 micron filter
- For questions on this certificate of analysis please contact techincal service at aagam444@gmail.com or +91 97247 47919

Computer generated Report - Does Not Require Signature

## **R.D. BROTHERS**